

CERTIFICATE OF ANALYSIS

PRODUCT CT0023B

FD10 FUSIDIC ACID

LOT NUMBER

1291939

EXPIRY DATE

2016.01.31

DATE OF

MANUFACTURE 2013.01.07

Delivery/Customer information

Date Printed

2013.11.08

Delivery No.

Customer

Customer Order number

Microbiological Performance

Assay (mcg/disc)

9.9

Tested in accordance with current BSAC methodology

	Zone Size (mm)	Limits (mm)
Staphylococcus aureus ATCC®9144	34	32 - 40
Staphylococcus aureus ATCC®25923	32	30 - 37

The above have been tested in our Quality Control Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. The results reported were obtained at the time of release.

Lot Accepted. 2013.01.09

Carol MacBride

PRODUCT PERFORMANCE MANAGER

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

