



CERTIFICATE OF ANALYSIS

Page 1 of 1

PRODUCT PO5091A
CAMPYLOBACTER CCDA SELECTIVE MEDIUM

LOT NUMBER 1301923

EXPIRY DATE 2014.08.29

TYPICAL FORMULATION

	grams per litre
'Lab-Lemco' powder	10.00000
Peptone	10.00000
Sodium chloride	5.00000
Bacteriological Charcoal	4.00000
Casein hydrolysate	3.00000
Sodium desoxycholate	1.00000
Ferrous sulphate	0.25000
Sodium pyruvate	0.25000
Cefoperazone	0.03200
Amphotericin B	0.01000
Agar	12.00000

General Characteristics	Results	Specification
Colour	Conforms	Jet black
Appearance	Conforms	Opaque
pH	7.5	7.2 - 7.6
Packaging/Presentation	Conforms	Conforms to specification
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Campylobact. jejuni ssp.jejuni ATCC®33291	55	52	2-4mm, flat, grey-green colonies
Escherichia coli ATCC®25922	1E+05	0	Complete inhibition (<= 10 cfu)
Staphylococcus aureus ATCC®25923	1E+05	0	Complete inhibition (<= 10 cfu)
Enterococcus faecalis ATCC®29212	1E+05	0	Complete inhibition (<= 10 cfu)
Batch homogeneity acc ISO11133	<10000	Conforms	C.coli ATCC33559 test ag previous batch

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature