



CERTIFICATE OF ANALYSIS

PRODUCT MT0397
MOELLER DECARB LYSINE 1ML 13x100MM
13 X 100 MM

LOT NUMBER 1306

EXPIRY DATE 03.05.2012
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

Customer Order number

| Physical Characteristics | Results | Specification |
|------------------------------------|---------------|---------------------------|
| Appearance | Purple-yellow | Conforms to Specification |
| pH | 6.00 | 5.80 - 6.20 |
| Sterility @ 35 °C ± 2°C for 3 days | Conforms | Conforms to specification |

| Microbiological Performance Organism Tested | Expected Results | Specification |
|--|------------------------------------|---------------------------|
| Enterobacter cloacae ATCC® 35030 | Negative (yellow) | Conforms to Specification |
| Enterobacter aerogenes ATCC® 3048 | Positive (purple to purple yellow) | Conforms to Specification |
| Proteus vulgaris ATCC® 33420 | Negative (yellow) | Conforms to Specification |
| Pseudomonas aeruginosa ATCC® 27853 | Negative (unchanged-yellow brown) | Conforms to Specification |
| Shewanella putrifaciens ATCC® 49138 | Negative (unchanged-yellow brown) | Conforms to Specification |
| Burkholderia cepacia ATCC® 25416 | Positive (purple) | Conforms to Specification |

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager