



## CERTIFICATE OF ANALYSIS

**PRODUCT** MT2040  
TODD HEWITT BRW/5% SHEEP BLOOD,GENT,NAL  
5ML 16X100MM

**LOT NUMBER** 1306984

**EXPIRY DATE** 03.04.2013  
(dd.mm.yyyy)

### Delivery/Customer information

Date Printed (dd.mm.yyyy)

08.11.2013

Delivery No.

Customer

106343

ZDELETE Alpha Supplies

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Clear pale amber broth wi	Conforms to Specification
pH	7.70	7.60 - 8.00
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Inhibition	Conforms to Specification
Streptococcus agalactiae ATCC® 12386	Growth,complet/part.lysis of blood cells	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Inhibition	Conforms to Specification
Streptococcus pyogenes ATCC® 19615	Growth,complet/part.lysis of blood cells	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell  
Quality Control Manager