



CERTIFICATE OF ANALYSIS

Page 1 of 1

PRODUCT PO5079C
MALT EXTRACT AGAR

LOT NUMBER 1384149

EXPIRY DATE 2013.12.27

TYPICAL FORMULATION

	grams per litre
Malt extract	30.00000
Mycological peptone	5.00000
Agar	18.00000

General Characteristics	Results	Specification
Colour	Conforms	Ochre yellow
Appearance	Conforms	Transparent
pH	5.5	5.2 - 5.6
Packaging/Presentation	Conforms	Conforms to specification
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Candida albicans ATCC®10231	55	56	2-3mm, cream colonies
Aspergillus brasiliensis ATCC®16404	38	34	10-30mm, white mycelium, black spores

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature