



CERTIFICATE OF ANALYSIS

PRODUCT **MP1450**
NEW YORK CITY MEDIUM W/LCAT SUPPLEMENT

LOT NUMBER 1401863

EXPIRY DATE 12.11.2013
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

08.11.2013

Delivery No.

Customer

106343

ZDELETE Alpha Supplies

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Clear red/brown	Conforms to Specification
pH	7.60	7.20 - 7.60
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Inhibition (partial)	Conforms to Specification
Proteus mirabilis ATCC® 43071	Inhibition (partial)	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Inhibition (partial)	Conforms to Specification
Staphylococcus epidermidis ATCC® 12228	Inhibition (partial)	Conforms to Specification
Neisseria gonorrhoeae ATCC® 43069	Growth	Conforms to Specification
Neisseria meningitidis ATCC® 13090	Growth	Conforms to Specification
Neisseria sicca ATCC® 9913	Inhibition (partial)	Conforms to Specification
Candida albicans ATCC® 60193	Inhibition (partial)	Conforms to Specification
Neisseria gonorrhoeae ATCC® 43070	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager