



# CERTIFICATE OF ANALYSIS

**PRODUCT** CT0453B  
CFR30 CEFADROXIL

**LOT NUMBER** 1414117

**EXPIRY DATE** 2016.11.30

**DATE OF MANUFACTURE** 2013.10.25

Delivery/Customer information
Date Printed 2013.11.08
Delivery No.
Customer
Customer Order number

## Microbiological Performance

Assay (mcg/disc) 28.75

### Tested in accordance with current SRGA methodology

	Zone Size (mm)	Limits (mm)
<i>Escherichia coli</i> ATCC@25922	17	16 - 22

The above have been tested in our Quality Control Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. The results reported were obtained at the time of release.

Lot Accepted. 2013.10.29

Carol MacBride  
PRODUCT PERFORMANCE MANAGER

ATCC is a registered trade mark of the American Type Culture Collection.  
NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



**OXOID LIMITED**  
Wade Road, Basingstoke, Hampshire RG24 8PW. England  
oxid@thermofisher.com www.oxid.com