

CERTIFICATE OF ANALYSIS

PRODUCT MT1316

STAPH BROTH MODIFIED 8.5 ML 100/PK

16 X 100MM

LOT NUMBER

1415114

EXPIRY DATE (dd.mm.yyyy)

26.02.2014

26.02.2014

Delivery/Customer information

Date Printed (dd.mm.yyyy)

08.11.2013 Delivery No.

Customer 106343

ZDELETE Alpha Supplies

Customer Order number

Physical Characteristics	Results	Specification
Appearance pH Sterility @ 35 °C ± 2°C for 3 days	Clear straw colored 7.10 Conforms	Conforms to Specification 6.80 - 7.20 Conforms to specification

Microbiological Performance Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Inhibition	Conforms to Specification
Staphylococcus aureus ATCC® 43300	Growth	Conforms to Specification
Staphylococcus aureus ATCC® 29213	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell

Quality Control Manager

Køren Campbell