



CERTIFICATE OF ANALYSIS

PRODUCT **MP1315**
 ESBL ISOLATION AGAR

LOT NUMBER 1440415

EXPIRY DATE 10.02.2014
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

05.06.2014

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Pink	Conforms to Specification
pH	7.10	6.90 - 7.30
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance

Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Inhibition	Conforms to Specification
Escherichia coli ATCC® 51446	Growth of pink to red colonies	Conforms to Specification
Klebsiella pneumoniae ATCC® 700603	Growth of mucoid pink to red colonies	Conforms to Specification
Klebsiella pneumoniae ATCC® 13883	Inhibition	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager