



CERTIFICATE OF ANALYSIS

PRODUCT MT1953
SIMMON'S CITRATE AGAR SLANT 7ML 16x125MM

LOT NUMBER 1446169

EXPIRY DATE 08.07.2014
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

05.06.2014

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Green Slant	Conforms to Specification
pH	7.10	6.80 - 7.20
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Citrate- No Growth No colour change	Conforms to Specification
Klebsiella pneumoniae ATCC® 13883	Citrate+Growth with colour change	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager