

**CERTIFICATE OF ANALYSIS**

PRODUCT CT0183B
FOS50 FOSFOMYCIN

LOT NUMBER 1449842

EXPIRY DATE 2017.02.28

**DATE OF
MANUFACTURE** 2014.01.16

Delivery/Customer information

Date Printed

2014.06.05

Delivery No.

Customer

Customer Order number

Microbiological Performance

Assay (mcg/disc) 59.5

Tested in accordance with current SFM methodology

	Zone Size (mm)	Limits (mm)
<i>Staphylococcus aureus</i> ATCC®25923	27	24 - 35

The above have been tested in our Quality Control Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. The results reported were obtained at the time of release.

Lot Accepted. 2014.01.31

Carol MacBride

PRODUCT PERFORMANCE MANAGER

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

**OXOID LIMITED****Wade Road, Basingstoke, Hampshire RG24 8PW. England****oxoid@thermofisher.com www.oxoid.com**

FDA Reg No. 8010096