



## CERTIFICATE OF ANALYSIS

**PRODUCT**            **MP2012**  
**TRYPTONE SOYA AGAR W/5% SHEEP BLOOD**

**LOT NUMBER**            1454963

**EXPIRY DATE**            21.03.2014  
(dd.mm.yyyy)

### Delivery/Customer information

Date Printed (dd.mm.yyyy)

05.06.2014

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Cherry Red	Conforms to Specification
pH	7.30	7.10 - 7.50
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

### Microbiological Performance

Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Growth	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Growth	Conforms to Specification
Streptococcus pyogenes ATCC® 19615	Growth, beta haemolysis, neg. CAMP reaction	Conforms to Specification
Streptococcus pneumoniae ATCC® 6305	Growth, alpha haemolysis	Conforms to Specification
Streptococcus galactiae ATCC® 12386	Growth, positive CAMP reaction	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell  
Quality Control Manager