



CERTIFICATE OF ANALYSIS

PRODUCT **MP1573**
OXACILLIN RESISTENT SCREENING AGAR ORSAB

LOT NUMBER 1458695

EXPIRY DATE 28.03.2014
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

05.06.2014

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Clear blue/grey	Conforms to Specification
pH	7.00	7.00 - 7.40
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance	Expected Results	Specification
Organism Tested		
Proteus vulgaris ATCC® 33420	Inhibition	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Inhibition	Conforms to Specification
Staphylococcus aureus ATCC® 29213	Inhibition	Conforms to Specification
Staphylococcus aureus ATCC® 33591	Growth	Conforms to Specification
Staphylococcus aureus ATCC® 43300	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager