



CERTIFICATE OF ANALYSIS

PRODUCT **MT2009**
TETRATHIONATE BROTH 9ML 16x125 100 PCK

LOT NUMBER 1489880

EXPIRY DATE 14.10.2014
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

05.06.2014

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Cloudy white broth	Conforms to Specification
pH	8.20	7.80 - 8.20
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance	Expected Results	Specification
Organism Tested		
Salmonella typhimurium ATCC® 14028	Growth	Conforms to Specification
Escherichia coli ATCC® 25922	Inhibition, partial to complete	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager