



CERTIFICATE OF ANALYSIS

PRODUCT **MP1350**
 MUELLER HINTON AGAR

LOT NUMBER 1503549

EXPIRY DATE 12.08.2014
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

05.06.2014

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Clear amber	Conforms to Specification
pH	7.30	7.20 - 7.40
Depth of media	4.2	3.5 - 5.0
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance

Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Growth	Conforms to Specification
Escherichia coli ATCC® 35218	Growth	Conforms to Specification
Pseudomonas aeruginosa ATCC® 27853	Growth	Conforms to Specification
Enterococcus faecalis ATCC® 33186	Growth	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Growth	Conforms to Specification
Haemophilus influenzae ATCC® 10211	Growth, XV+, V-, X-	Conforms to Specification
Haemophilus parainfluenzae ATCC® 7901	Growth XV+, V+, X-	Conforms to Specification
Staphylococcus aureus ATCC® 43300	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager