

**CERTIFICATE OF ANALYSIS**

PRODUCT CM0325B
PLATE COUNT AGAR 500g

LOT NUMBER 1741087

EXPIRY DATE 2020.09.28

DATE OF MANUFACTURE 2015.09.15

Delivery/Customer information

Date Printed
 2015.10.01
 Delivery No.

Customer
 Customer Order number

Physical Characteristics	Results	Specification
Appearance	Straw powder	Straw powder
Colour on reconstitution	Straw 1-2	Straw 1-2
pH (25°C)	7.0	6.8 - 7.2
Clarity	Clear	Clear
Thermophiles	Absent	Absent
Mesophiles	Absent	Absent

Microbiological Performance	Control cfu	Test cfu	Recovery %	Test Description
Aerobic incubation at 30°C for 48 hours				
Pour plate technique				
<i>Escherichia coli</i> ATCC®25922	24	31	129	Straw cols
<i>Escherichia coli</i> ATCC®8739	40	45	113	Straw cols
<i>Staphylococcus aureus</i> ATCC®6538	52	54	104	Straw cols
<i>Staphylococcus aureus</i> ATCC®6538P	97	97	100	Straw cols
<i>Bacillus subtilis</i> ATCC®6633	11	14	127	Straw cols

Control Medium: Tryptone Soya Agar

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium from an inoculum of 10-100 colony-forming units (cfu).

All ISO/CEN 11133-2 control strains are included in the test panel.

Refer to product specification for full details.



Tested by the Product Performance Laboratory
 OXOID LIMITED


Wade Road, Basingstoke, Hampshire RG24 8PW. England
 oxoid@thermofisher.com www.oxoid.com



CERTIFICATE OF ANALYSIS

PRODUCT CM0325B
PLATE COUNT AGAR 500g
LOT NUMBER 1741087

The information given is believed to be correct. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Product Performance Laboratory, Oxoid Limited, Basingstoke. The results reported were obtained at the time of release.
Lot Accepted. 2015.09.23


.....
Carol MacBride
Product Performance Manager

ATCC is a registered trade mark of the American Type Culture Collection.
NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



Tested by the Product Performance Laboratory
OXOID LIMITED
Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxoid@thermofisher.com www.oxoid.com