

**CERTIFICATE OF ANALYSIS**

PRODUCT CM0509R
BUFFERED PEPTONE WATER 2.5 kg

LOT NUMBER 1804932

EXPIRY DATE 2021.01.31

DATE OF MANUFACTURE 2016.01.14

Delivery/Customer information

Date Printed
 2016.01.19

Delivery No.

Customer
 Customer Order number

Physical Characteristics	Results	Specification
Appearance	Straw powder	Straw powder
Colour on reconstitution	Straw 2	Straw 1-2
pH (25°C)	7.3	7.0 - 7.4
Clarity	Clear	Clear

Microbiological Performance	Initial Inoculum	Test Result
-----------------------------	------------------	-------------

Tested as a non-selective pre-enrichment broth**Aerobic incubation at 37°C for 18 hours**

<i>Salmonella typhimurium</i> ATCC®14028	28	Turbid growth
<i>Salmonella enteritidis</i> ATCC®13076	22	Turbid growth
<i>Salmonella nottingham</i> NCTC7832	26	Turbid growth
<i>Escherichia coli</i> ATCC®11775	32	Turbid growth


Control Medium: Tryptone Soya Agar

A satisfactory result is represented by visible growth from an inoculum of 10-100 colony-forming units (cfu).
 All ISO/CEN 11133-2 control strains are included in the test panel.
 Refer to product specification for full details.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Product Performance Laboratory, Oxoid Limited, Basingstoke.

The results reported were obtained at the time of release.

Lot Accepted. 2016.01.18



 Carol MacBride
 Product Performance Manager

ATCC is a registered trade mark of the American Type Culture Collection.
 NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



Tested by the Product Performance Laboratory
OXOID LIMITED
 Wade Road, Basingstoke, Hampshire RG24 8PW. England
 oxid@thermofisher.com www.oxid.com

FDA Reg No. 8010096