



CERTIFICATE OF ANALYSIS

PRODUCT **MBB0009**
1% THIOGLYCOLIC ACID 100 ML
125 ML CLEAR ICHEM CR00586

LOT NUMBER 2016

EXPIRY DATE 17.07.2012
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

08.11.2013

Delivery No.

Customer

106343

ZDELETE Alpha Supplies

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Clear, hazy, faint pink/p	Conforms to Specification

Microbiological Performance Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Growth	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Growth	Conforms to Specification
Bacillus subtilis ATCC® 6633	Growth	Conforms to Specification
Pseudomonas aeruginosa ATCC® 27853	Growth	Conforms to Specification
Clostridium perfringens ATCC® 13124	Growth	Conforms to Specification
Candida albicans ATCC® 10231	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager