



CERTIFICATE OF ANALYSIS

PRODUCT **MT2059**
TRYPTONE SOYA BROTH 2 ML 13x100MM PCK100

LOT NUMBER 2025

EXPIRY DATE 26.07.2012
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

08.11.2013

Delivery No.

Customer

106343

ZDELETE Alpha Supplies

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Pale amber broth	Conforms to Specification
pH	7.20	7.10 - 7.50
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Growth	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Growth	Conforms to Specification
Streptococcus pneumoniae ATCC® 6305	Growth	Conforms to Specification
Streptococcus pyogenes ATCC® 19615	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager