



## CERTIFICATE OF ANALYSIS

**PRODUCT** MT1564  
O-F MEDIUM W/ MANNITOL 3ML 13x100MM

**LOT NUMBER** 2104

**EXPIRY DATE** 13.10.2012  
(dd.mm.yyyy)

### Delivery/Customer information

Date Printed (dd.mm.yyyy)

08.11.2013

Delivery No.

Customer

106343

ZDELETE Alpha Supplies

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Green stab	Conforms to Specification
pH	6.80	6.60 - 7.00
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Burkholderia cepacia ATCC® 25416	Oxidation	Conforms to Specification
Acinetobacter lwoffii ATCC® 17925	No change	Conforms to Specification
Klebsiella pneumoniae ATCC® 13883	Fermentation	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell  
Quality Control Manager