



CERTIFICATE OF ANALYSIS

PRODUCT MT1650
PHOS. BUFFER SALINE (pH 7.6 M1/15) 13MM
5 ML 13 X 100 MM

LOT NUMBER 2193

EXPIRY DATE 10.01.2013
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

08.11.2013

Delivery No.

Customer

106343

ZDELETE Alpha Supplies

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Clear liquid	Conforms to Specification
pH	7.60	7.40 - 7.80
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager