

CERTIFICATE OF ANALYSIS

PRODUCT MT0350

COOKED MEAT MEDIUM 10 ML.16x125MM

LOT NUMBER

2256

EXPIRY DATE

14.03.2013

(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

08.11.2013

Delivery No.

Customer 106343

ZDELETE Alpha Supplies

Customer Order number

Physical Characteristics	Results	Specification	
Appearance pH	Clear amber 7.00	Conforms to Specification 7.00 - 7.40	
Sterility @ $35 ^{\circ}\text{C} \pm 2 ^{\circ}\text{C}$ for 3days	Conforms	Conforms to specification	

Microbiological Performance Organism Tested	Expected Results	Specification
Staphylococcus aureus ATCC® 25923	Growth	Conforms to Specification
Bacteroides fragilis ATCC® 25285	Growth	Conforms to Specification
Clostridium perfringens ATCC® 13124	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell

Quality Control Manager

Køren Campbell