



## CERTIFICATE OF ANALYSIS

**PRODUCT** MT1815  
BORDETELLA TRANSPORT SEL 5ML16x100MM

**LOT NUMBER** 2271

**EXPIRY DATE** 22.11.2012  
(dd.mm.yyyy)

### Delivery/Customer information

Date Printed (dd.mm.yyyy)

08.11.2013

Delivery No.

Customer

106343

ZDELETE Alpha Supplies

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Black stab	Conforms to Specification
pH	7.50	7.20 - 7.60
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Inhibition	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Inhibition	Conforms to Specification
Bordetella bronchiseptica ATCC® 10580	Growth	Conforms to Specification
Bordetella pertussis ATCC® 12742	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell  
Quality Control Manager