

Fisher Diagnostics a division of Fisher Scientific Company, LLC a part of Thermo Fisher Scientific Inc. 8365 Valley Pike Middletown, VA (USA) 22645-1905 Tel: 540 869 3200 or 800-528-0494

Certificate of Analysis

Product Name: Contour Normal Ctrl Product Number: L1001698 Lot Number: 2352101 Expiration Date: 2013-10-31 Storage: Room Temp

Specification	Specification Limit	Result
Appearance	Clear, red solution free of particulate	Complies
	matter	
Specific Gravity @ 25 ± 0.2 °C	1.0029 - 1.0051	1.0048
Glucose Content	121.5 - 126.5 mg/dL	124.4

For in-vitro diagnostic use.

The above lot number was manufactured in compliance with Quality Systems Regulations, has met all product specifications and has been released.

Thiland

Quality Representative

Date: 2011-10-27