

Fisher Diagnostics a division of Fisher Scientific Company, LLC a part of Thermo Fisher Scientific Inc. 8365 Valley Pike Middletown VA (USA) 22645-1905

Middletown, VA (USA) 22645-1905 Tel: 540 869 3200 or 800-528-0494

## Certificate of Analysis

Product Name: Contour Normal Single Ctrl Expiration Date: 2014-01-31
Product Number: 06707504 Storage: Room Temp

Lot Number: 2382012

Specification	Specification Limit	Result
Appearance	Clear, red solution free of particulate	Complies
	matter	
Specific Gravity @ 25 ± 0.2 °C	1.0029 - 1.0051	1.0048
Glucose Content	121.5 - 126.5 mg/dL	123.0

For in-vitro diagnostic use.

The above lot number was manufactured in compliance with Quality Systems Regulations, has met all product specifications and has been released.

Quality Representative

Connie Carroll

Date: 2012-01-31

PRINTED ON: 2013-11-11