

*Certificate of Analysis*


Product Name: Ascensia Contour Low Ctrl  
Product Number: 1859  
Lot Number: 2597122

Expiration Date: 2014-12-31  
Storage: Room Temp

Specification	Specification Limit	Result
Appearance	Clear, red solution free of particulate matter	Complies
Specific Gravity @ 25 ± 0.2 °C	1.0025 - 1.0038	1.0036
Glucose Content	43.5 - 46.5 mg/dL	44.6

For in-vitro diagnostic use.

The above lot number was manufactured in compliance with Quality Systems Regulations, has met all product specifications and has been released.



Quality Representative

Date: 2013-01-06