

Certificate of Analysis

Product Name: High Purity Water (5 gallons)
Product Numbers: 9805
Lot Number: 267161

Production Date: 14-DEC-12
Expiration Date: 2013-12-31
Storage: 15 - 30°C

Specification	Specification Limit	Result
Silicates	<0.05 PPM	0.00
Microbial Content	<10 CFU/mL	0
Resistivity*	>=10 Mohmcm	19
Conductivity*	<2.0 µmhos/cm	0.9
Total Organic Carbon*	<0.50 mg C/L	0.12

For in-vitro diagnostic use. For laboratory use.

* At time of manufacture.

High Purity Water (CAP/CLSI CLRW)

Prepared to meet C.L.S.I (formerly NCCLS) Standards for Clinical Laboratory Reagent Water (CLRW) at the time of manufacture. Prepared typically at 18 megohm-cm resistance using reverse osmosis, mixed bed deionization, activated carbon and filtration to 0.2 micron. Not intended for irrigation, injection or infusion. Contains no preservatives.



Quality Representative

Date: 2012-12-21