

Fisher Diagnostics a division of Fisher Scientific Company, LLC a part of Thermo Fisher Scientific Inc. 8365 Valley Pike

Middletown, VA (USA) 22645-1905 Tel: 540 869 3200 or 800-528-0494

Certificate of Analysis

Product Name: Contour Next Normal Control Expiration Date: 2014-12-31 Product Number: 84168662 Storage: Room Temp

Lot Number: 2709063

Specification	Specification Limit	Result
Appearance	Clear, red solution free of	Complies
	particulates.	
Specific Gravity @ 25 ± 0.2 °C	1.0017 - 1.0027	1.0021
Glucose Content	65.2 - 68.8 mg/dL	66.4

For in-vitro diagnostic use.

The above lot number was manufactured in compliance with Quality Systems Regulations, has met all product specifications and has been released.

Quality Representative

Connie Carroll Date: 2013-07-08

PRINTED ON: 2013-11-08