remel	

Certificate of Quality

Product Name: E coli 0157:H7 VT (N) Non-Tox NCTC 12900 Lot Number: 641566

Product Number: C1967L Expiration Date: 2009-09-30 (YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Purity:

Standardized aliquots of the rehydrated product are inoculated onto nonselective media and examined for pure growth following the appropriate incubation. Selective and Differential media are also tested where applicable.

Viability And Quantification:

Each organism is recovered from the preserved state within the required time frame and at an acceptable level. Passage number is stated as the current preserved state.

Macroscopic And Microscopic Morphology:

Colony morphology is consistent with documented referenced description. Traditional staining is performed.

Biochemical Analysis:

Organism exhibits characteristic biochemical and/or enzymatic reactions. Automated and/or conventional testing was performed and results were within established limits. Antimicrobial testing performed where applicable. Results within expected ranges.

CFU/ loop: >10(4) Passage: 3 Gram Reaction: Gram Negative Rods Biochemical Profile: Vitek GNI + & Biochemical testing E.coli serogroup O157 positive / H7 negative from initial growth E.coli H7 positive after passage through motility media to enhance motility. Appearance: Preserved Gel Matrix suspended in loop. pH: N/A

Signed

Quality Control Supervisor