



CERTIFICATE OF ANALYSIS

Page 1 of 1

PRODUCT PB5105A
ENTEROHAEMOLYSIN AGAR WITH BLOOD

LOT NUMBER 683256

EXPIRY DATE 2008.09.26

TYPICAL FORMULATION

	grams per litre
Columbia Blood Agar Base	39.00000
Calcium chloride	1.47000
Washed Sheep erythrocytes	40.00000 ML

General Characteristics	Results	Specification
Colour	Conforms	Traffic red
Appearance	Conforms	Opaque
pH	7.0	6.9 - 7.3
Packaging/Presentation	Conforms	Conforms to specification
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Escherichia coli 01 :K1 DSM10830	<10000	Conforms	Good growth, grey colonies
Escherichia coli ATCC®25922	<10000	Conforms	Good growth, cream cols with haemolysis
Escherichia coli DSM13526	<10000	Conforms	Good growth, grey cols, enterohaemolysin+

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature