



# CERTIFICATE OF ANALYSIS

Page 1 of 1

**PRODUCT** PO5004A  
NEISSERIA SELECTIVE MEDIUM PLUS

**LOT NUMBER** 783845

**EXPIRY DATE** 2009.08.10

## TYPICAL FORMULATION

	grams per litre
Special peptone	15.00000
Corn starch	1.00000
Sodium chloride	5.00000
Dipotassium hydrogen phosphate	4.00000
Potassium dihydrogen phosphate	1.00000
Haemoglobin	10.00000
Vitox	20.00000 ML
Vancomycin	0.00500
Colistin methane sulphate	0.00750
Amphotericin B	0.00800
Trimethoprim	0.00500
Agar	10.00000

General Characteristics	Results	Specification
Colour	Conforms	Signal brown
Appearance	Conforms	Opaque
pH	7.0	7.0 - 7.4
Packaging/Presentation	Conforms	Conforms to specification
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Neisseria gonorrhoeae ATCC®49226	28	33	2-3 mm, brown colonies
Enterococcus faecalis ATCC®29212	5.0E+04	0	Complete inhibition (<= 10 cfu)
Candida albicans ATCC®10231	1.9E+04	0	Complete inhibition (<= 10 cfu)

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature