



CERTIFICATE OF ANALYSIS

PRODUCT PO5068A
DST AGAR
For in vitro diagnostic use.

LOT NUMBER 797954

EXPIRY DATE 2010.01.18

TYPICAL FORMULATION

	grams per litre
Proteose peptone	10.00000
Veal infusion solids	10.00000
Glucose	2.00000
Sodium chloride	3.00000
Disodium phosphate	2.00000
Sodium acetate	1.00000
Adenine sulphate	0.01000
Guanine hydrochloride	0.01000
Uracil	0.01000
Xanthine	0.01000
Aneurine	0.00002
Agar	12.00000

General Characteristics	Results	Specification
Colour	Conforms	Ivory
Appearance	Conforms	Transparent
pH	7.2	7.2 - 7.6
Packaging/Presentation	Conforms	Conforms to specification
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	Within Limits

Susceptibility Testing	Results(mm)	Limits (mm)
Escherichia coli ATCC®35218 - AMC30	19	16 -22
Escherichia coli ATCC®35218 - CTX30	39	36 -43
E.coli ATCC®35218 - TOB10	23	18 -26
E. coli ATCC®35218 - CIP5	36	36 -41
Pseudom. aeruginosa ATCC®27853 - ATM30	30	28 -33
Ps. aeruginosa ATCC®27853 - CAZ10	28	28 -36
Pseudom. aeruginosa ATCC®27853 - IPM10	24	20 -25
Ps. aeruginosa ATCC®27853 - PRL30	30	26 -32
S. aureus ATCC®25923 - OX5	35	30 -40
S. aureus ATCC®25923 - E15	33	28 -37
S. aureus ATCC®25923 - VA30	22	20 -26
S. aureus ATCC®25923 - KZ30	40	35 -45
Enterococcus faecalis ATCC®29212 - F100	26	25 -29
Enterococcus faecalis ATCC®29212 - AMP10	31	26 -32
Enterococcus faecalis ATCC®29212 - CIP5	24	21 -27

The information given is believed to be correct. However, both the information and the product are offered without warranty



CERTIFICATE OF ANALYSIS

Page 2 of 2

for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature