



## CERTIFICATE OF ANALYSIS

**PRODUCT**            **MP0805**  
                              **HAEMOPHILUS TEST MEDIUM**

**LOT NUMBER**            8168

**EXPIRY DATE**            08.09.2008  
(dd.mm.yyyy)

### Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

106343

ZDELETE Alpha Supplies

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Light amber	Conforms to Specification
pH	7.30	7.20 - 7.60
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

### Microbiological Performance

Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 35218	growth	Conforms to Specification
Haemophilus influenzae ATCC® 10211	growth	Conforms to Specification
Haemophilus influenzae ATCC® 49247	growth	Conforms to Specification
Haemophilus influenzae ATCC® 49766	growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell  
Quality Control Manager