



CERTIFICATE OF ANALYSIS

PRODUCT **MP1209**
 LEGIONELLA BCYE MED w/ BMPA

LOT NUMBER 8193

EXPIRY DATE 03.10.2008
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

106343

ZDELETE Alpha Supplies

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Black	Conforms to Specification
pH	6.90	6.70 - 7.10
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Inhibition	Conforms to Specification
Staphylococcus epidermidis ATCC® 12228	Inhibition	Conforms to Specification
Legionella pneumophila ATCC® 33152	Good growth; grey white-blue colonies	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager