



CERTIFICATE OF ANALYSIS

PRODUCT **MP0330**
 CHOCOLATE ENRICHED

LOT NUMBER 8326

EXPIRY DATE 13.02.2009
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

106343

ZDELETE Alpha Supplies

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Chocolate brown	Conforms to Specification
pH	7.20	7.00 - 7.40
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Neisseria meningitidis ATCC® 13090	Growth	Conforms to Specification
Neisseria gonorrhoeae ATCC® 43069	Growth	Conforms to Specification
Haemophilus parainfluenzae ATCC® 7901	Growth	Conforms to Specification
Haemophilus influenzae ATCC® 49766	Growth	Conforms to Specification
Haemophilus influenzae ATCC® 10211	Growth	Conforms to Specification
Neisseria gonorrhoeae ATCC® 43070	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager