



## CERTIFICATE OF ANALYSIS

**PRODUCT**            **MP1353**  
**MUELLER HINTON AGAR W/5% SHEEP BLOOD 150**

**LOT NUMBER**            8343

**EXPIRY DATE**            02.02.2009  
(dd.mm.yyyy)

### Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

106343

ZDELETE Alpha Supplies

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Cherry Red	Conforms to Specification
pH	7.40	7.20 - 7.40
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Streptococcus pneumoniae ATCC® 49619	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell  
Quality Control Manager