



CERTIFICATE OF ANALYSIS

Page 1 of 1

PRODUCT PO5041A
CAMPYLOBACTER KARMALI SELECTIVE MEDIUM

LOT NUMBER 834653

EXPIRY DATE 2010.02.05

TYPICAL FORMULATION

	grams per litre
Columbia Blood Agar Base	39.00000
Activated charcoal	4.00000
Haemin	0.03200
Sodium pyruvate	0.10000
Cefoperazone	0.03200
Vancomycin	0.02000
Amphotericin B	0.01000

General Characteristics	Results	Specification
Colour	Conforms	Jet black
Appearance	Conforms	Opaque
pH	7.4	7.2 - 7.6
Packaging/Presentation	Conforms	Conforms to specification
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Campylobact. jejuni ssp.jejuni ATCC@33291	92	97	2-3mm, grey shiny colonies
Escherichia coli ATCC@25922	2.6E+04	0	Complete inhibition (<= 10 cfu)
Staphylococcus epidermidis ATCC@12228	5.8E+04	0	Complete inhibition (<= 10 cfu)
Campylobacter coli ATCC@43478	<10000	Conforms	Good growth of characteristic colonies

For positive strains, colony count is greater than or equal to 50% of the control medium.
Tested in accordance with CEN ISO/TS 11133-2.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature