



CERTIFICATE OF ANALYSIS

PRODUCT **MP2260**
 VOGEL JOHNSON AGAR

LOT NUMBER 9086

EXPIRY DATE 08.05.2009
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

106343

ZDELETE Alpha Supplies

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Orange	Conforms to Specification
pH	7.10	7.00 - 7.40
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Inhibition (marked to complete)	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Out of specification	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager