



## CERTIFICATE OF ANALYSIS

**PRODUCT** MT2130  
UREA AGAR SLANT (CHRISTENSEN) 3ML PCK100  
13X100MM

**LOT NUMBER** 9120

**EXPIRY DATE** 30.10.2009  
(dd.mm.yyyy)

### Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

106343

ZDELETE Alpha Supplies

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Pale orange slant	Conforms to Specification
pH	6.70	6.60 - 7.00
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

### Microbiological Performance

Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Urea negative, no colour change	Conforms to Specification
Proteus vulgaris ATCC® 33420	Urea positive, pink slant 3 to 5 hours	Conforms to Specification
Klebsiella pneumoniae ATCC® 13883	Weak positive, pink slant after 18 hours	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell  
Quality Control Manager