



CERTIFICATE OF ANALYSIS

PRODUCT **MP1560**
SORBITOL MACCONKEY AGAR

LOT NUMBER 9125

EXPIRY DATE 28.07.2009
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

106343

ZDELETE Alpha Supplies

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Pink	Conforms to Specification
pH	7.12	6.90 - 7.30
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Growth, pink colonies	Conforms to Specification
Enterococcus faecalis ATCC® 29212	Inhibition	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Inhibition	Conforms to Specification
Proteus mirabilis ATCC® 43071	Growth:colorless col.swarming inhibited	Conforms to Specification
Escherichia coli O157:H7 ATCC® 43888	Growth, colorless colonies	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager