



## CERTIFICATE OF ANALYSIS

**PRODUCT**            **MP0425**  
                         **DC MEDIUM W/BCIG**

**LOT NUMBER**            9232

**EXPIRY DATE**            15.10.2009  
(dd.mm.yyyy)

### Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

106343

ZDELETE Alpha Supplies

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Clear orange/red	Conforms to Specification
pH	7.40	7.00 - 7.40
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

### Microbiological Performance

Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Growth, purple-blue colonies	Conforms to Specification
Salmonella typhimurium ATCC® 14028	Growth, clear colonies	Conforms to Specification
Enterococcus faecalis ATCC® 29212	Inhibition (Partial to complete)	Conforms to Specification
Citrobacter freundii ATCC® 8090	Growth, pink-purple colonies	Conforms to Specification
Enterobacter cloacae ATCC® 35030	Growth, pink colonies	Conforms to Specification
Klebsiella pneumoniae ATCC® 13883	Growth, pink mucoid colonies	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Inhibition (Partial to complete)	Conforms to Specification
Pseudomonas aeruginosa ATCC® 27853	Inhibition (Partial to complete)	Conforms to Specification
Aeromonas hydrophila ATCC® 35654	Inhibition (Partial to complete)	Conforms to Specification
Pseudomonas putida ATCC® 49128	Inhibition (Partial to complete)	Conforms to Specification
Escherichia coli ATCC® 8739	Growth, spd col ;pk edges ;purp-blu cent	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell  
Quality Control Manager