



CERTIFICATE OF ANALYSIS

PRODUCT MT0286
BLEB W/MOPS 9.9 ML 16X125 MM PCK 100

LOT NUMBER 9257

EXPIRY DATE 16.03.2010
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

106343

ZDELETE Alpha Supplies

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Straw coloured	Conforms to Specification
pH	7.30	7.10 - 7.50
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Listeria monocytogenes ATCC® 7646	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager