

*Certificate of Analysis*

Product Name: 100L Liquid Corr Test  
Product Numbers: SJ4M034E99  
Lot Number: 957947  
Millipore PO No.: 1165527  
P-Spec No./Rev. SJ4M034E99PU / C


Date of Manufacture: 10/08/2010  
Expiration Date: 2013-10-08  
Quantity: 2

Specification	Specification Limit	Result
Interfacial Tension	0.489 - 0.527 Dyn/cm	0.523
Density - Intrusion Fluid @ 22 °C	1.10 - 1.12 g/cm <sup>3</sup>	1.11
Density - Wetting Fluid @ 22 °C	1.07 - 1.09 g/cm <sup>3</sup>	1.08

For in-vitro diagnostic use.

The above lot number has been evaluated by the Quality Assurance department and has met all acceptance criteria as defined by Millipore finished product specification.

The above lot will be retained at least until the date of expiration. This product is approved and has been released.



Quality Representative      Date: 2010-10-19