

*Certificate of Analysis*

Product Name: 8 Liter Liquid Corr Test  
Product Numbers: 40502  
Lot Number: 969245  
Millipore PO No.: 1167099  
P-Spec No./Rev. 40502PU / C

Date of Manufacture: 11-NOV-10  
Expiration Date: 2013-11-11  
Quantity: 34

Specification	Specification Limit	Result
Interfacial Tension	0.489 - 0.527 Dyn/cm	0.520
Density - Intrusion Fluid @ 22 °C	1.10 - 1.12 g/cm <sup>3</sup>	1.11
Density - Wetting Fluid @ 22 °C	1.07 - 1.09 g/cm <sup>3</sup>	1.08

For in-vitro diagnostic use.

The above lot number has been evaluated by the Quality Assurance department and has met all acceptance criteria as defined by Millipore finished product specification.

The above lot will be retained at least until the date of expiration. This product is approved and has been released.



Quality Representative

Date: 2010-11-19