



# CERTIFICATE OF ANALYSIS

Page 1 of 1

**PRODUCT** PB5034A  
SCHAEDLER + SHEEP BL + HAEM + VIT.K1

**LOT NUMBER** 975820

**EXPIRY DATE** 2011.02.04

## TYPICAL FORMULATION

	grams per litre
Tryptone soya broth	10.00000
Special peptone	5.00000
Yeast extract	5.00000
Glucose	5.00000
Sodium chloride	4.00000
Cysteine HCl	0.40000
Tris buffer	0.75000
Haemin	0.01000
Vitamine K1	0.00200
Agar	13.50000
Defibrinated sheep blood	50.00000 ML

General Characteristics	Results	Specification
Colour	Conforms	Signal red
Appearance	Conforms	Opaque
pH	7.5	7.4 - 7.8
Packaging/Presentation	Conforms	Conforms to specification
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Bacteroides fragilis ATCC@25285	71	68	2-4mm, white shiny colonies
Peptostreptococcus anaerobius ATCC@27337	66	42	3-5mm, white shiny colonies
Clostridium perfringens ATCC@13124	<10000	Conforms	Good growth, cream cols, double zone haem

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature