

Fisher Diagnostics a division of Fisher Scientific Company, LLC a part of Thermo Fisher Scientific Inc. 8365 Valley Pike Middletown, VA (USA) 22645-1905 Tel: 540 869 3200 or 800-528-0494

Certificate of Analysis

Product Name: TRIGLYCERIDE, 10X50ML Product Numbers: TR22203 Lot Number: 998462 Date of Manufacture: 17-JAN-11 Expiration Date: 2013-03-31 Storage: 2 - 8°C

| Specification | Specification Limit | Result |
|----------------------------|----------------------------------------|-------------|
| Appearance (Powder) | White to off-white free-flowing powder | Complies |
| Appearance (Reconstituted) | Clear, slightly yellow solution, | Complies |
| | essentially | |
| | free of particulate matter | |
| Data-Trol Normal | Lower Limit | 0.70 mmol/L |
| Data-Trol Normal | Upper Limit | 1.11 mmol/L |
| Data-Trol Normal | Result | 0.88 mmol/L |
| Data-Trol Normal | Lot Number | V36526 |
| Data-Trol Abnormal | Lower Limit | 1.26 mmol/L |
| Data-Trol Abnormal | Upper Limit | 1.67 mmol/L |
| Data-Trol Abnormal | Result | 1.41 mmol/L |
| Data-Trol Abnormal | Lot Number | V36528 |
| Linearity | Linear to 11.5 mMol/L | Complies |

For in-vitro diagnostic use.

The above lot number was manufactured in compliance with Quality Systems Regulations, has met all product specifications and has been released.

Quality Representative

Date: 2011-04-08