

Certificate of Analysis

Product Name: V-Lyte Standard B/Salt Bridge
Product Numbers: K825
Lot Number: KB3A01

Production Date: 2013-01-09
Expiration Date: 2014-01-02

Description	Part Number	Component Lot Number
Standard B	JCK825B	258452
Salt Bridge	JCK825S	258448

Specification	Specification Limit	Result
Standard B		
Appearance	Clear, no precipitates	Complies
Sodium	9.78 - 10.08 mmol/L	9.93
Potassium	0.198 - 0.202 mmol/L	0.200
Chloride	5.83 - 6.03 mmol/L	5.89
pH at 25 °C	9.30 - 9.50	9.38
Ave. Dissolved Oxygen Conc. In-Process	<3.5 PPM	2.6
Ave. Dissolved Oxygen Conc. End Use	<6.5 PPM	4.1
Air Volume (cc)	<3.5	0.2

Salt Bridge		
Appearance	Clear, no precipitates	Complies
Sodium	<0.100 mmol/L	0.021
Potassium	118.0 - 122.0 mmol/L	119.9
Chloride	118.33 - 122.33 mmol/L	120.32
pH at 25 °C	3.5 - 3.9	3.7
Ave. Dissolved Oxygen Conc. In-Process	<3.5 PPM	1.9
Ave. Dissolved Oxygen Conc. End Use	<6.5 PPM	4.5
Air Volume (cc)	<3.5	0.1

For in-vitro diagnostic use.



Quality Representative

Date: 2013-01-15