



CERTIFICATE OF ANALYSIS

PRODUCT **MP1337**
 MARTIN LEWIS AGAR W/LCAT

LOT NUMBER 1059

EXPIRY DATE 25.04.2011
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Chocolate brown	Conforms to Specification
pH	7.30	7.00 - 7.40
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance	Expected Results	Specification
Organism Tested		
Escherichia coli ATCC® 25922	Inhibition (partial)	Conforms to Specification
Proteus mirabilis ATCC® 43071	Inhibition (partial)	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Inhibition (partial)	Conforms to Specification
Staphylococcus epidermidis ATCC® 12228	Inhibition (partial)	Conforms to Specification
Neisseria gonorrhoeae ATCC® 43069	Growth	Conforms to Specification
Neisseria meningitidis ATCC® 13090	Growth	Conforms to Specification
Candida albicans ATCC® 60193	Inhibition (partial)	Conforms to Specification
Neisseria sicca ATCC® 9913	Inhibition (partial)	Conforms to Specification
Neisseria gonorrhoeae ATCC® 43070	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager