



CERTIFICATE OF ANALYSIS

PRODUCT **PO0169A**
BILE AESCULIN AGAR
1 PACK OF 10 PLATES

LOT NUMBER 1079816

EXPIRY DATE 2009.04.23

Delivery/Customer information
Date Printed 2013.11.11
Delivery No.
Customer
Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Straw/brown
pH (25°C)	7.3	6.9 - 7.3
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Enterobacter aerogenes ATCC®13048	42	23	Brown cols, aesculin hydrolysis
Streptococcus pyogenes ATCC®19615	>10000	0	No growth
Enterococcus faecalis ATCC®29212	42	40	Brown cols, aesculin hydrolysis

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd