



CERTIFICATE OF ANALYSIS

PRODUCT **PB0308A**
CNA STAPH/STREP SELECT. AGAR (MODIFIED)
1 PACK OF 10 PLATES

LOT NUMBER 1099102

EXPIRY DATE 2009.09.15

Delivery/Customer information	
Date Printed	2013.11.11
Delivery No.	
Customer	
Customer Order number	

Physical Characteristics	Results	Specification
Appearance	Conforms	Red opaque
pH (25°C)	7.5	7.1 - 7.5
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Streptococcus pyogenes ATCC@19615	86	64	Pale straw cols, beta haemolysis
Staphylococcus aureus ATCC@25923	48	37	cream colonies
Proteus mirabilis NCTC10975	>10000	0	± growth no swarming
Pseudomonas aeruginosa ATCC@27853	>10000	0	No Growth

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd